

Contact Resistivity – Contact Spot

- Finger to finger resistance measurements
- Resistance range : 0.1 – 1000 ohms
- Accuracy : 0.4%
- Minimum probe pitch: 1.3 mm (every other finger for smaller pitches)

